

## Measurement Uncertainty Data for Quantitative Food Microbiology Tests

PathWest Method Number	Analyte	Method Type	Units	Result	Lower limit	Upper limit	% Variation
MFH001	Total Plate Count	Pour Plate	CFU per gram	100	90	111	11%
MFH010	<i>Bacillus cereus</i>	Spread Plate	CFU per gram	100	77	129	26%
MFH014	Coag Pos <i>Staphylococci</i>	Spread Plate	CFU per gram	100	79	126	24%
MFH034	<i>E.coli</i>	Pour Plate	CFU per gram	100	81	124	22%
MFH034	Total Coliforms	Pour Plate	CFU per gram	100	94	106	6%
MFH045	Total Plate Count	Petrifilm®	CFU per gram	100	77	130	27%
MFH046	<i>E.coli</i>	Petrifilm®	CFU per gram	100	83	120	19%

Last updated: 23-Apr-10