

Measurement Uncertainty Data For Quantitative Water Microbiology Tests

PathWest Method Number	Analyte	Method Type	Units	Result	Lower limit	Upper limit	% Variation
MWM001	Total Coliforms	Membrane Filtration	CFU per 100mL	100	82	121	20%
MWM002	Thermotolerant Coliforms	Membrane Filtration	CFU per 100mL	100	85	117	16%
MWM002	<i>E.coli</i>	Membrane Filtration	CFU per 100mL	100	81	123	21%
MWM006	Standard Plate Count	Pour Plate	CFU per mL	100	84	119	18%
MWM013	<i>Legionella</i> spp.	Spread Plate	CFU per mL	100	84	118	17%
MWM022	Heterotrophic Plate Count	Pour Plate	CFU per mL	100	93	108	8%

Last updated: 23-Apr-10